



RAE Systems, Inc.  
1339 Moffett Park Dr.  
Sunnyvale, CA 94089 USA

Telephone 408-752-0723  
Tube Orders 1-888-723-8823  
Web www.raesystems.com

# Gas Detection Tube Data Sheet

## Trichloroethylene CHCl=CCl<sub>2</sub> No. 10-119-20

	Extended Range	Standard Range	Extended Range
Range (ppmv)	2.5 - 50	5 - 100	10 - 230
No. of Pump Strokes	2	1	0.5
Sample Volume (mL)	200	100	50
Sample Time (min)	2 x 3	3	2
Correction Factor	0.5	1	2.3

Precision (Relative Standard Deviation)\*:  $\leq \pm 20\%$

Linearity with No. of Pump Strokes:  $r^2 = 0.999$

Humidity: No effect 0 - 95% RH

Temperature Range: 0 - 40°C  
(32 - 104°F)

Temp (°C/°F)	0/32	10/50	25/77	40/104
Corr. Factor	1.6	1.3	1.0	1.1

Storage Life and Conditions: 1 year in darkness at 5 - 25°C (40 - 77°F)

Color Change: Yellow → Purple

Reaction Principle: Cl<sub>2</sub>C=CHCl + PbO<sub>2</sub> + H<sub>2</sub>SO<sub>4</sub> → HCl  
HCl + Base → Chloride (dye color change)

<u>Cross-sensitivity:</u> Substance	Concentration (ppmv)	Apparent Reading*
Tetrachloroethylene	40	70
1,2-Dichloroethylene	100	20
Vinyl Chloride	100	10
1,1,2-Trichloroethane	100	<0.5
Acetone	1000	0
Toluene	1000	0
p-Xylene	1000	0
Cl <sub>2</sub>	10	10 (pale beige)
HCl	50	21
NO	500	0
NO <sub>2</sub>	500	60 (pale beige)

\*Data based on RAE pumps and tubes used in standard range.

Other Possible Interferences: Acid gases. No response to H<sub>2</sub>S, CO or CH<sub>4</sub>.

**Caution: Use of connector tubing other than that supplied may reduce response.**

Caution: Dispose of spent or expired tubes according to local regulations.  
Possibly hazardous materials are given under the section Reaction Principle.